

Amendments to the Claims:

1. (Currently Amended) A system for interacting with targeted tissue of a patient and for use in conjunction with a remote programmer and a power source, the system comprising:

a central control module adapted for subcutaneous implantation, the central control module comprising:

a housing configured to house the power source; and

a wireless receiver disposed in said housing, the wireless receiver configured to be coupled to the power source and to receive programming signals from the remote programmer;

a first lead coupled to the central control module;

a satellite module adapted for subcutaneous implantation coupled to said central control module by means of said first lead, said satellite module comprising:

a processor coupled to said wireless receiver and configured to be coupled to the power source;

a communication module coupled to said processor for communicating with said central control module;

a switching module coupled to said processor;

a memory coupled to said communication module and said processor;

a sense amp; and

an A/D converter coupled to said sense amp and said memory; and

a plurality of second leads adapted for subcutaneous implantation coupled to said sense amp and said switching module, each of said second leads having at least one tissue interaction element,

wherein said processor is configured to receive said programming signals from said wireless receiver and to control said switching module ~~is configured, in response to said processor,~~ to selectively distribute or receive signals to or from the plurality of second leads.

2.-3. (Cancelled)

4. (Currently Amended) The system for interacting with targeted tissue of a patient as recited in claim 1[[3]] wherein the central control module further includes a communication module coupled to the power source and the satellite module wherein communication between the central control module and the satellite module is combined with power delivery to minimize interconnections.
5. (Original) The system for interacting with targeted tissue of a patient as recited in claim 4 wherein the central control module is limited to providing power to the satellite module and acting as a communication hub of the system.
6. (Original) The system for interacting with targeted tissue of a patient as recited in claim 1 wherein the satellite module includes a signal generator coupled to the switching module for generating a therapy stimulation signals for the targeted tissue.
7. (Original) The system for interacting with targeted tissue of a patient as recited in claim 1 wherein the satellite module further includes:
 - a logic block coupled to the switching module; and
 - a buffer coupled to the logic block and the memory.
8. (Original) The system for interacting with targeted tissue of a patient as recited in claim 1 wherein the central control module and satellite module is limited to two wire interconnections for providing power and communication.
9. (Original) The system for interacting with targeted tissue of a patient as recited in claim 1 wherein the central control module and satellite module is limited to four wire interconnections for providing power, communication, and stimulation signals.
10. (Original) The system for interacting with targeted tissue of a patient as recited in claim 1 wherein control of a therapy stimulation signals to the targeted tissue of a patient is controlled by the satellite module.

11. (Withdrawn) A system for interacting with targeted tissue of a patient, the system comprising:

- a central control module; and

- a satellite module coupled to the central control module placed in proximity to the targeted tissue wherein the satellite module includes:

- a hybrid circuit substrate;

- at least one integrated circuit mounted on an upper surface of the hybrid circuit substrate;

- at least one via through the hybrid circuit substrate wherein the at least one integrated circuit couples to the at least one via at the upper surface of the hybrid circuit substrate;

- at least one lead adapted for subcutaneous implantation coupled to said at least one via at a bottom surface of the hybrid circuit substrate, the at least one lead having at least one tissue interaction element; and

- a cover hermetically sealed to an upper surface of the hybrid circuit substrate covering the at least one integrated circuit and the at least one via.

12. (Withdrawn) The system for interacting with targeted tissue of a patient as recited in claim 11 further including at least one contact on the upper surface of the hybrid circuit substrate wherein a portion of the at least one contact is outside the cover and wherein the at least one contact is coupled to the at least one integrated circuit.

13. (Withdrawn) The system for interacting with targeted tissue of a patient as recited in claim 12 wherein the hybrid circuit substrate comprises a ceramic material.

14. (Withdrawn) The system for interacting with targeted tissue of a patient as recited in claim 13 further including a dielectric layer between the cover and the at least one contact and wherein the at least one contact couples to a lead adapted for subcutaneous implantation having at least one tissue interaction element.

15. (Withdrawn) The system for interacting with targeted tissue of a patient as recited in claim 14 wherein wiring between the central control module and the satellite module is limited to two wires and wherein the two wires are used to provide power from the central control module to the satellite module and bi-directional communication between the central control module and the satellite module.

16. (Withdrawn) The system for interacting with targeted tissue of a patient as recited in claim 14 wherein wiring between the central control module and the satellite module is limited to four wires or less.

17. (Withdrawn) A method of communicating between a central control module and a satellite module for interacting with targeted tissue of a patient comprising the steps of:

- providing a programming word comprising information to be sent to the satellite module;

- converting the programming word into identifiable groups of pulses corresponding to bits of the programming word;

- sending the identifiable groups of pulses to the satellite module;

- converting the identifiable groups of pulses in the satellite module to information for providing a therapy to the patient; and

- storing the energy from the identifiable groups of pulses to power the satellite module.

18. (Withdrawn) The method of communicating between a central control module and a satellite module for interacting with targeted tissue of a patient as recited in claim 17 wherein the step of providing a programming word comprising information to be sent to the satellite module further including the steps of:

- beginning the programming word with a plurality of bits identifying a start of the programming word;

- following the plurality of bits identifying the start of the programming word with a plurality of information bits; and

- ending the programming word with a cyclic redundancy code (CRC).

19. (Withdrawn) The method of communicating between a central control module and a satellite module for interacting with targeted tissue of a patient as recited in claim 18 wherein the step of following the plurality of bits identifying the start of the programming word with a plurality of information bits further including the steps of:

- providing a plurality of bits corresponding to a module address; and
- providing a plurality of bits corresponding to electrode settings.

20. (Withdrawn) The method of communicating between a central control module and a satellite module for interacting with targeted tissue of a patient as recited in claim 19 further including the steps of:

- receiving the cyclic redundancy code at the satellite module;
- applying a polynomial to the received information to generate a data check word;
- comparing the cyclic redundancy code against the data check word; and
- identifying to the central control module whether the information was sent correctly.

21. (Withdrawn) The method of communicating between a central control module and a satellite module for interacting with targeted tissue of a patient as recited in claim 17 further including the steps of:

- providing X bits of the programming word;
- reading the X bits as a corresponding decimal equivalent;
- sending a number of pulses equal to the corresponding decimal equivalent;
- providing a predetermined delay; and

repeating said steps of providing X bits of the programming word, reading the X bits as a decimal equivalent, sending a number of pulses equal to the corresponding decimal equivalent, and providing a predetermined delay until all bits of the programming word have been converted and sent.

22. (Withdrawn) A method of operating a system for interacting with targeted tissue of a patient, the system includes a central control module and a satellite module both adapted for subcutaneous implantation, the satellite module being placed in proximity to the targeted tissue, the method comprising the steps of:

- sensing analog signals from electrodes coupled to the targeted tissue;
- amplifying the analog signals in the satellite module;
- converting the analog signals to digital words;
- storing the digital words in memory on the satellite module;
- processing the digital words in the satellite module;
- generating a therapy from analysis of the digital words in the satellite module; and
- implementing the therapy.

23. (Withdrawn) The method of operating a system for interacting with targeted tissue of a patient, the system includes a central control module and a satellite module both adapted for subcutaneous implantation, the satellite module being placed in proximity to the targeted tissue as recited in claim 22 further including the steps of:

- storing in memory information received from more than one electrode;
- sending the information from the more than one electrode to the central control module;

and

- sending the information from the more than one electrode by radio frequency transmission from the central control module to a programmer.

24. (Withdrawn) The method of operating a system for interacting with targeted tissue of a patient, the system includes a central control module and a satellite module both adapted for subcutaneous implantation, the satellite module being placed in proximity to the targeted tissue as recited in claim 23 further including the steps of:

- analyzing the information;
- sending a modification to the therapy by radio frequency transmission from the programmer to the central control module;

- sending the modification to the therapy from the central control module to the satellite module;

- storing the modification to the therapy in the satellite module; and
- implementing the modification to the therapy.

25. (New) An implantable medical system for use in conjunction with a remote programmer and a power source, the implantable medical system comprising:

- a central control module, comprising:

 - a housing configured to support the power source; and

 - a wireless receiver disposed in the housing, the wireless receiver configured to be coupled to the power source and to receive programming signals from the remote programmer;

- an elongated conductor having a distal end portion and a proximal end portion, the proximal end portion coupled to the central control module;

- a plurality of tissue-interactive elements; and

- a satellite control module coupled between the plurality of tissue-interactive elements and the distal end portion of the elongated conductor, the satellite control module configured to selectively route electrical signals to selected ones of the plurality of tissue-interactive elements in accordance with the programming signals received from the remote programmer.